

EC Declaration of Conformity

Manufacturer: HeartSine Technologies Limited
207 Airport Road West
Belfast, BT3 9ED
Northern Ireland
United Kingdom

Device: **samaritan® PAD 360P**
Model: **SAM 360P**
Description: Automated external defibrillator.
Medical Device Classification: Identified as Class IIb under rule 9 of Annex IX of Council Directive 93/42/EEC as amended by 2007/47/EC

Scope of Declaration for Australia: Refer to Appendix 2.

HeartSine Technologies declares that the HeartSine samaritan PAD 360P (SAM 360P), a therapeutic medical device in the range of Automated External Defibrillators, and its associated accessories are designed and manufactured in conformity with:

- a) The essential requirements (Annex I) and provisions of the European Medical Device Directive (MDD) European Council Directive 93/42/EEC (as amended by 2007/47/EC)
 - And is subject to the procedure set out in Annex II (excluding section 4), Full Quality Assurance System, of Directive 93/42/EEC, as amended by Directive 2007/47/EC;
 - Under the supervision of TÜV SÜD Product Service GmbH, (Notified Body Number 0123), TÜV SÜD Product Service GmbH, Certification Body, Ridlerstraße 65, 80339 Munich, Germany.
- b) Clause 1.8 of Schedule 3 of the Australian Therapeutic Goods (Medical Devices) Regulations 2002.
 - Each kind of medical device to which the system has been applied complies with the applicable provisions of the essential principles, classification rules, and the full quality assurance procedures, at each stage, from the design of the device until its final inspection before being supplied.
 - It is subject to the Australian Standards Applied referred within Appendix 3.
- c) ROHS Directive (2011/65/EU), amended by RoHS3 Directive (EU 2015/863), with exemptions Annex IV number 17 lead solder in portable defibrillators, Annex III exemption 6c – copper alloy containing up to 4% lead by weight, exemption 7(a) – lead in high melting solders, exemption 7 (c)-I - Electrical and electronic components containing lead in glass or ceramics.
- d) HeartSine Technologies European Authorised Representative address is as follows; Stryker European Operations Limited, Anngrove, IDA Business & Technology Park, Carrigtwohill, Co Cork, T45HX08, Ireland.

HeartSine Technologies is exclusively responsible for this declaration of conformity.

Certification

Council Directive 93/42/EEC
EN ISO 13485 : 2016

TÜV Certificate Number

No. G1 067590 0006 Rev. 01
No. Q5 067590 0008 Rev. 00

Signature



*Electronically signed by:
Rebecca Funston
Reason: I approve
Date: Apr 21, 2021 13:05
GMT+1*

Date

21-Apr-2021**Rebecca Funston**

**Director, Global Regulatory & Clinical Affairs
HeartSine Technologies Ltd.**

Appendix 1

Ref	Description	GMDN Code
360-STR-CA-10	1 SAM 360P with 1 Adult Pad-Pak	48047
360-STR-CZ-10	1 SAM 360P with 1 Adult Pad-Pak	48047
360-STR-DA-10	1 SAM 360P with 1 Adult Pad-Pak	48047
360-STR-DE-10	1 SAM 360P with 1 Adult Pad-Pak	48047
360-STR-EE-10	1 SAM 360P with 1 Adult Pad-Pak	48047
360-STR-EL-10	1 SAM 360P with 1 Adult Pad-Pak	48047
360-STR-ES-10	1 SAM 360P with 1 Adult Pad-Pak	48047
360-STR-FI-10	1 SAM 360P with 1 Adult Pad-Pak	48047
360-STR-FR-10	1 SAM 360P with 1 Adult Pad-Pak	48047
360-STR-HR-10	1 SAM 360P with 1 Adult Pad-Pak	48047
360-STR-HU-10	1 SAM 360P with 1 Adult Pad-Pak	48047
360-STR-IS-10	1 SAM 360P with 1 Adult Pad-Pak	48047
360-STR-IT-10	1 SAM 360P with 1 Adult Pad-Pak	48047
360-STR-LT-10	1 SAM 360P with 1 Adult Pad-Pak	48047
360-STR-LV-10	1 SAM 360P with 1 Adult Pad-Pak	48047
360-STR-NL-10	1 SAM 360P with 1 Adult Pad-Pak	48047
360-STR-NO-10	1 SAM 360P with 1 Adult Pad-Pak	48047
360-STR-PO-10	1 SAM 360P with 1 Adult Pad-Pak	48047
360-STR-PT-10	1 SAM 360P with 1 Adult Pad-Pak	48047
360-STR-RK-10	1 SAM 360P with 1 Adult Pad-Pak	48047
360-STR-RO-10	1 SAM 360P with 1 Adult Pad-Pak	48047
360-STR-SL-10	1 SAM 360P with 1 Adult Pad-Pak	48047
360-STR-ST-10	1 SAM 360P with 1 Adult Pad-Pak	48047
360-STR-SV-10	1 SAM 360P with 1 Adult Pad-Pak	48047
360-STR-TR-10	1 SAM 360P with 1 Adult Pad-Pak	48047
360-STR-UK-10	1 SAM 360P with 1 Adult Pad-Pak	48047
360-STR-CA-AV	1 SAM 360P with 1 Aviation Pad-Pak	48047
360-STR-CZ-AV	1 SAM 360P with 1 Aviation Pad-Pak	48047
360-STR-DA-AV	1 SAM 360P with 1 Aviation Pad-Pak	48047
360-STR-DE-AV	1 SAM 360P with 1 Aviation Pad-Pak	48047
360-STR-EE-AV	1 SAM 360P with 1 Aviation Pad-Pak	48047
360-STR-EL-AV	1 SAM 360P with 1 Aviation Pad-Pak	48047
360-STR-ES-AV	1 SAM 360P with 1 Aviation Pad-Pak	48047
360-STR-FI-AV	1 SAM 360P with 1 Aviation Pad-Pak	48047
360-STR-FR-AV	1 SAM 360P with 1 Aviation Pad-Pak	48047
360-STR-HR-AV	1 SAM 360P with 1 Aviation Pad-Pak	48047
360-STR-HU-AV	1 SAM 360P with 1 Aviation Pad-Pak	48047
360-STR-IS-AV	1 SAM 360P with 1 Aviation Pad-Pak	48047
360-STR-IT-AV	1 SAM 360P with 1 Aviation Pad-Pak	48047
360-STR-LT-AV	1 SAM 360P with 1 Aviation Pad-Pak	48047
360-STR-LV-AV	1 SAM 360P with 1 Aviation Pad-Pak	48047
360-STR-NL-AV	1 SAM 360P with 1 Aviation Pad-Pak	48047
360-STR-NO-AV	1 SAM 360P with 1 Aviation Pad-Pak	48047
360-STR-PO-AV	1 SAM 360P with 1 Aviation Pad-Pak	48047
360-STR-PT-AV	1 SAM 360P with 1 Aviation Pad-Pak	48047
360-STR-RK-AV	1 SAM 360P with 1 Aviation Pad-Pak	48047
360-STR-RO-AV	1 SAM 360P with 1 Aviation Pad-Pak	48047
360-STR-SL-AV	1 SAM 360P with 1 Aviation Pad-Pak	48047
360-STR-SV-AV	1 SAM 360P with 1 Aviation Pad-Pak	48047

360-STR-TR-AV	1 SAM 360P with 1 Aviation Pad-Pak	48047
360-STR-UK-AV	1 SAM 360P with 1 Aviation Pad-Pak	48047

Appendix 2

Model

Item Number	Description	AU GMDN Code
360-BAS-UK-10	SAM 360P 1 Pad-Pak-03 UK English	48047
360-TSO-UK-10	SAM 360P 1 Pad-Pak-07 UK English	48047
360-BAS-USROW-10	SAM 360P 1 Pad-Pak-03 US English	48047
360-TSO-USROW-10	SAM 360P 1 Pad-Pak-07 US English	48047
360-BAS-SJ-10	SAM 360P 1 Pad-Pak-03 US English	48047
360-BAS-UK-GW	PACKAGE, 360P, GATEWAY	48047
360-BAS-UR-GW	PACKAGE, 360P, GATEWAY	48047
360-STR-UK-10	SAM 360P, Pad-Pak-03, UK English	48047
360-STR-UK-AV	SAM 360P, Pad-Pak-07, UK English	48047

Appendix 3

Standards	Title
ISO 13485	Medical devices – Quality management systems – Requirements for regulatory purposes
AAMI MDS	Developing safe, effective and reliable medical software
ANSI/AAMI EC57	Testing and reporting performance results of cardiac rhythm and ST segment measurement algorithms
EN 1041	Requirements for information supplied by medical device manufacturers
ISO 60529	Degrees of protection provided by enclosures (IP Code)
ISO 15223	Symbols to be used with medical device labels, labelling and information to be supplied
ISO 14971	Application of risk management to medical devices
ISO 10993	Biological evaluation of medical devices – Part 1: Evaluation and testing within a risk management process
IEC 60878	Graphical Symbols for Electrical Equipment in Medical Practice
21 CFR 820	Quality System Regulation
EN 60601-1	General requirements for safety for medical electrical equipment
EN 60601-1-2	EMC requirements for medical electrical equipment
EN 60601-1-6	Safety requirements for usability
EN 60601-2-4	Safety requirements for cardiac defibrillators